

TECHNICAL BULLETIN

#46 January 5, 1989

IMPROVED FACEPIECE MOUNTED REGULATOR

SURVIVAIR has replaced the facepiece mounted regulator P/N 910504 with a new improved assembly P/N 940160. The new regulator uses most of the old parts, but four additional parts allow flow adjustments to be made from the outside while the regulator is operating.

This is accomplished as follows:

- a) Remove adjustment plug.
- b) Use the SURVIVAIR test bench – P/N 905700.
- c) During flow test (test #4) use 1/8" allen wrench to adjust flow.
- d) When flow is within tolerance, reinstall adjustment plug.

All other procedures remain the same.

